

Medaval Comparative-Equivalence Assessment Full Report

Volume 2024

Report 2403FR

06 February 2024

Full report on the comparative-equivalence assessment of the blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 and the SELVAS ACCUNIQ BP500 upper arm blood pressure monitors, according to the requirements of MEDDEV 2.7/1 revision 4.

Analysis Neil Atkins, Medaval Ltd., Dublin, IRELAND.

Administration Con Creedon Medaval Ltd., Dublin, IRELAND.

Reference Medaval Ltd. Medical Full report on the comparative-equivalence assessment of the blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 and the SELVAS ACCUNIQ BP500 upper arm blood pressure monitors, according to the requirements of MEDDEV 2.7/1 revision 4. *Medical Device Assessment*. 2024 Feb 06;2024(2403FR) 34 p.

Medical Device Assessment is published by
Medaval Ltd., Unit 107, SBC, Serpentine Ave., Ballsbridge, Dublin D04 H522, IRELAND.

© 2024 Medaval Ltd. All rights reserved.

Permissions: Requests for permissions to reproduce figures, tables, or portions of reports or articles originally published in *Medical Device Assessment* can be obtained by email request to info@medaval.ie.

MEDICAL DEVICE ASSESSMENT 2403FR:2024

Full report on the comparative-equivalence assessment of the blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 and the SELVAS ACCUNIQ BP500 upper arm blood pressure monitors, according to the requirements of MEDDEV 2.7/1 revision 4.

Medaval Comparative-Equivalence Assessment Full Report – 06 February 2024

Contents

GLOSSARY	3
ABBREVIATIONS	3
LEGEND OF ITEM COMPARISON	3
LEGEND OF SUMMARY COMPARISON	4
SUMMARY	5
ORGANISATIONAL DETAILS	6
MEDAVAL LTD.....	6
SELVAS HEALTHCARE INC.....	6
FULL REPORT	8
INTRODUCTION.....	8
METHODOLOGY	8
SELVAS ACCUNIQ BP500 v SELVAS ACCUNIQ BP503 RESULTS	8
CUFF COMPARISON RESULTS	9
SELVAS ACCUNIQ BP500 v SELVAS ACCUNIQ BP503 DISCUSSION	9
CUFF COMPARISON DISCUSSION	9
CONCLUSION	10
RECOMMENDATIONS	10
REFERENCES	10
SELVAS ACCUNIQ BP503 V BP500 COMPARATIVE-EQUIVALENCE	12
SUMMARY OF DIFFERENCES	12
COMPARISON OF PRIMARY DETAILS	12
COMPARISON OF STANDARD DEVICE DETAILS	14
COMPARISON OF BPM-SPECIFIC DETAILS	19
SUMMATION OF COMPARISONS	26
EQUIVALENCE AND COMPARATIVE ANALYSIS.....	27
CUFF COMPARATIVE-EQUIVALENCE	28
SUMMARY OF DIFFERENCES	28
COMPARISON OF DETAILS	28
SUMMATION OF COMPARISONS	30
EQUIVALENCE AND COMPARATIVE ANALYSIS.....	31
APPENDIX 1 – MEDAVAL COMPARATIVE-EQUIVALENCE PROCEDURE	32
COVER PAGE.....	32
SUMMARY	33
APPENDIX 2 –COMPARATIVE-EQUIVALENCE OF ALL DEVICES IN FAMILY	34

Glossary

Abbreviations

%RH	relative humidity percent
°C	degrees Celsius
AAMI	Association for the Advancement of Medical Instrumentation (USA)
ABPM	Ambulatory blood pressure measurement
ANSI	American National Standards Institute
BP	Blood pressure
bpm	beats per minute
cm	centimetre(s)
DBP	Diastolic blood pressure
ECG	Electrocardiogram
EEA	European Economic Area (EU plus Iceland, Liechtenstein and Norway)
EEC	European Economic Community (former name for EU)
ESH	European Society of Hypertension
EU	European Union
g	gram(s)
hPa	hectopascals
ISO	International Organization for Standardization/International Standards Organization
m	metre(s)
MAP	Mean Arterial Pressure (if estimated, $MAP = (SBP + 2 \times DBP) / 3$)
meas.	Measurement
min	minute(s)
mmHg	millimetre(s) of mercury
PC	Personal computer (any external system to which data can be downloaded)
PP	Pulse Pressure ($PP = SBP - DBP$)
PR	Pulse Rate
PRP	Pressure Rate Product ($PRP = SBP \times PR$)
req.	requirement
SBP	Systolic Blood Pressure
SD	standard deviation

Legend of Item Comparison

BL	Similar-level provisions on both devices
BP	Different provisions on each device
EL	Equivalent provision on both devices
EP	Identical provision on both devices
MB	Values missing for both devices
MR	Value missing for Reference Device
MT	Value missing for Test Device
QL	Check if provision is identical or similar
QP	Check free text entries
RL	Better provision on Reference Device
RP	Provided on Reference but not on Test Device
TL	Better provision on Test Device
TP	Provided on Test but not on Reference Device
XB	Not applicable for this device functionality
XP	Not comparable

Legend of Summary Comparison

An	Applicable number of items (total less not applicable)
Bn	Differing items that are not any better in either device
Bn-F	Differing feature items that are not any better in either device
Dn	Items that are different in the two devices
Dn-A	Accessory items that are different in the two devices
Dn-C	Core items that are different in the two devices
Dn-F	Feature items that are different in the two devices
En	Items that are equal in both devices
En-F	Core items that are equal in both devices
En-F	Feature items that are equal in both devices
Mn	Missing items
Mn-A	Missing accessory items
Mn-C	Missing core items
Mn-F	Missing feature items
Qn	Items with outstanding queries
Qn-A	Accessory items with outstanding queries
Qn-C	Core items with outstanding queries
Qn-F	Feature items with outstanding queries
Rn	Differing items that are better in the reference device than in the test device
Rn-F	Differing feature items that are better in the reference device than in the test device
Sn	Total number of items
Sn-A	Total number of accessory items
Sn-C	Total number of core items
Sn-F	Total number of feature items
Sn-I	Total number of identity items
Tn	Differing items that are better in the test device than in the reference device
Tn-F	Differing feature items that are better in the test device than in the reference device
Xn	Items that were on the checklist but were not compared
Xn-A	Accessory items that were on the checklist but were not compared
Xn-F	Feature items that were on the checklist but were not compared

Summary

Objective

The objective of this study was to compare the SELVAS ACCUNIQ BP503 blood pressure monitor (BPM) to the SELVAS ACCUNIQ BP500 for equivalence according to the requirements of the MEDDEV 2.7/1 revision 4 guidelines¹ in order to fulfil the medical device requirements of EU 2017/745².

Methodology

The 320 items in a checklist were compared, of which 72 were core items, to prove equivalence and broken down into 63 technical and nine clinical items. The 248 non-core items (not associated with the measurement technology) identified the differences between the devices and determined how the devices were comparable.

The cuffs, defined as clinical items in terms of the device, were analysed separately, as at least one difference was observable. The 16 items in the checklist consisted of ten core items, broken down nine technical items and one biological item. There were six non-core items.

Non-core items are grouped in into "identity", "feature" and "accessory" subgroups.

Results

Comparison of the core items, for both the devices and cuffs proved that they were equivalent according to procedures recommended in the

MEDDEV 2.7/1 revision 4 guidelines¹. Comparison of the non-core items, for both the devices demonstrated that, with fewer features and accessories, the SELVAS ACCUNIQ BP503 is considered simpler to the SELVAS ACCUNIQ BP500.

Both cuff styles were found to be identical, as regards blood pressure measurement, with the only difference being the colour of the front deco cover.

Conclusion

As they have been prove to be equivalent, the results on any validation carried out on either the SELVAS ACCUNIQ BP500 or the SELVAS ACCUNIQ BP503 must be applied to both devices.

As the SELVAS ACCUNIQ BP500 has been proven to be accurate according to the requirements of the ISO 81060-2:2018^{3,4} and ISO 81060-2:2018/Amd 1:2020 standard⁵, the results of this validation must be applied to the SELVAS ACCUNIQ BP503 also⁶. Therefore the SELVAS ACCUNIQ BP503 must be considered to be accurate, when used correctly, as per the manufacturer's instructions, within the criteria set out in ISO 81060-2:2018 and ISO 81060-2:2018/Amd 1:2020.

It is, therefore, recommended for use in clinical blood pressure measurement.

Organisational Details

Medaval Ltd.

Incorporated in 1989 as Medical Device Assessment Ltd, the company abbreviated its name to Medaval Ltd. in 2015. Medaval provides several services including comprehensive cardiovascular device listings according to peer-reviewed validations, certification for devices that have been proven to have been validated strictly according to a current standard protocol, validation of devices and comparative-equivalence according to MEDDEV 2.7/1 rev 4 standards¹. Both validation and comparative-equivalence services are in accordance with Regulation (EU) 2017/745².

The passing criteria in validation protocols are based on specific sample distributions and on other criteria and can only be applied if all of the requirements are followed correctly. Therefore, in any validation study, Medaval, first tests the hypothesis that the study was not carried out in accordance with the requirements and it is only if that hypothesis is rejected can the results be considered reliable.

All procedures were developed and reviewed by a panel of experts. The Medaval Accreditation Procedure is designed to check that every aspect of a validation protocol is fulfilled. Modifications, that may be necessary for particular populations or circumstances not defined specifically in a protocol must be supported by relevant peer-reviewed scientific publications.

Validation is considered to apply to the specific measurement technology being tested, as distinct from the device itself. No inference should be made about the validity of any other aspect of the device, unless it is also tested according to a regulatory or peer-reviewed protocol. Validation also only applies to the population from which the sample is taken and under the circumstances in which it was carried out, as defined in the protocol. No inference should be made about the validity of the device in a different population or under different circumstances.

The results must apply equally to any device that uses the same measurement technology, as proven under MEDDEV 2.7/1 rev 4 standards irrespective of whether that equivalence is proven prior to or subsequent to the validation. Medaval has developed as comparative-equivalence procedure to test the null hypothesis that two devices are not equivalent, according to this standard. Should that hypothesis be rejected, the devices must be regarded as equivalent for that measurement technology.

For more information, please refer to www.medaval.ie.

SELVAS Healthcare Inc.

SELVAS Healthcare Inc. (formerly Jawon Medical), established in 1993, is a digital healthcare company based on technology for medical devices and assistive rehabilitation technology devices.

The company has won several awards including the Presidential Award in the National Venture Awards and selected as a World Top-class Company (1999), Top Prize in the Leaders' Venture Awards (2000), the Prime Minister's Award by the Korean Good Manufacturing Practice Trade Day (2001), the Director's Award by the Korea Food and Drug Administration (2003), Bronze Prize in the Republic of Korea Technical Awards, Silver Prize in the Venture Design Awards and Bronze Medal of Industrial Effort in the Precision Technology Promotion Contest (2005), the Director's Award by the Korea Food and Drug Administration (2006), the Advanced Venture Company Award (2010), Grand Prize in the 1st People's Happiness Premium IT and the Popularity Award, Analysis and Diagnosis System Segment in the Korean Medical Device Awards (2014).

Its brand ACCUNIQ derives from its mission to provide accurate diagnosis and unique technology.

SELVAS Headquarters are in Daejeon, Republic of Korea and it has major offices in Beijing, Peoples Republic of China and in Austin, Texas, USA. It also has offices in

Venlo, Netherlands, Taipei City, Taiwan and Tokyo, Japan.

The company has, among several others, ISO 13485, ISO 9001 and GMP certification. See

<https://www.ACCUNIQ.com/en/company/certification.php> for a complete list.

For more information, please refer to www.ACCUNIQ.com.

Full Report

Introduction

Both the SELVAS ACCUNIQ BP500 and SELVAS ACCUNIQ BP503 are blood pressure monitors (BPMs) intended for waiting room use or for availability in, for instance, a staff common area in a business. Blood pressure (BP) is measured oscillometrically on each device in the same way. A visual aid shows how to place the, usually right, arm and a sensor prevents it from working until at least the elbow is placed correctly. Measurement is initiated by pressing a Start/Stop button and the results are displayed. It is intended for adults (18+) with arm circumferences in the range 20 cm to 40 cm only but not for those with arrhythmias and some other stated conditions.

An additional rear monitor, allowing anthropometric data to be entered is available and the device has a USB port and two RS-232C ports. None of these were used in this study.

This assessment is intended to test the null hypothesis that the oscillometric measurement technologies differ between the two devices.

Methodology

The devices were compared according to the requirements of the MEDDEV 2.7/1 revision 4 guidelines¹ and, in doing so, fulfil the requirements of EU 2017/745². According to these requirements, all aspects of the devices must be classified as being associated with technical, clinical or biological aspects the measurement technology being tested (core items) or as being associated with aspects other than the measurement technology.

With the input of several renowned experts in validation, Medaval drew up a procedure involving over 300 checks, with more, as necessary, depending on innovations on supplementary features and accessories. These are designed to fulfil the requirements of MEDDEV 2.7/1 revision 4¹, specified thereof in Appendix A1, Appendix A9 and Appendix A10. At the time of this study, there are 320

checks for BPMs and a further 16 checks for cuffs.

SELVAS ACCUNIQ BP500 v SELVAS ACCUNIQ BP503 Results

All of the requirements of the protocol were satisfied without any adjustments or violations.

For each equivalence check, there were 320 items to be compared, of which 72 were core items broken down into 63 technical and nine clinical items. While the cuffs are defined as clinical items in terms of the device, they are analysed separately and contain their own core items. The 248 non-core items were grouped into 11 "identity", 193 "feature" and 44 "accessory" subgroups.

The functionalities defined for 2 of the core items (analogue filter model and inflation target value) were not applicable to the device functionality of either of the devices and 14 which were not provided on either of the devices. This left 56 items, of which 55 were identical on both devices. The remaining item referred to the cuffs used. Separate equivalence procedures proved that the provision of these items were equivalent (the only difference being the colour of a deco cover on the front). Worth noting is that SELVAS uses one of two pumps in the production of these devices. Both pumps are manufactured specifically for the devices and details specification were provided to prove that they operate identically.

For the identity items, four were not applicable to either of the devices, five were identical on all devices, two (name and model number) were equivalent across both devices.

The functionalities defined for 77 of the feature items were not applicable to the device functionality of either of the devices, 68 were not provided on either of the devices and the remaining 48 were identical on both devices.

The functionalities defined for 15 of the accessory items were not applicable to the device functionality of either of the devices, ten were not provided on either of the devices and 18 were identical on both devices. The

difference in the remaining item related to the colours of the Start/Stop and the Emergency Stop buttons.

As information, including that of non-provision, was provided on all 320 items identified for comparison, the comparison and the identification of all differences, as defined by the protocol, was complete. The results of all comparisons are provided in summary and in detail below.

Cuff Comparison Results

All of the requirements of the protocol were satisfied without any adjustments or violations.

Two cuff designs are provided for which, in total, there were 16 items to be compared, of which 10 were core items broken down into nine technical items and one biological item. The six non-core items are grouped into four "identity", one "feature" and one "accessory" subgroups.

The functionality, relating to microphones, for one of the core items was not applicable and one not provided to either design and the remaining eight were identical on both designs. For the identity items, three were identical on both designs and one had equivalent provision. The feature item was identical on both designs. The single difference referred to the colour of the deco cover, a ring securing the front of the cuff to the casing).

As information, including that of non-provision, was provided on all 16 items identified for comparison, the comparison and the identification of all differences, as defined by the protocol, was complete. The results of all comparisons are provided in summary and in detail below.

SELVAS ACCUNIQ BP500 v SELVAS ACCUNIQ BP503 Discussion

The SELVAS ACCUNIQ BP503 and the SELVAS ACCUNIQ BP500 have been compared, with respect to oscillometric blood pressure measurement, according to a procedure verified as ensuring that the comparison has been conducted in accordance

with the requirements set out in MEDDEV 2.7/1 revision 4. Therefore, any hypothesis that the reliability of the results may be compromised due to protocol adjustment or violation must be rejected and the results must be considered to be valid.

According to this protocol, the results of the comparison require that the null hypothesis, that the monitors differ in respect to technical, clinical or biological core oscillometric blood pressure measurement technology in the two monitors, is rejected. Therefore, it must be concluded, that the oscillometric blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 is equivalent to that used in the SELVAS ACCUNIQ BP500. Furthermore, the results of any validations carried out this oscillometric blood pressure measurement technology, irrespective of which of the two monitors is used during the validation procedure, must be applied equally to the this technology of both monitors.

For more general use, SELVAS ACCUNIQ BP503 is considered identical to the SELVAS ACCUNIQ BP500, as there are no functional differences between these devices.

Cuff Comparison Discussion

The two styles of cuffs have been compared, with respect to oscillometric blood pressure measurement, according to a procedure verified as ensuring that the comparison has been conducted in accordance with the requirements set out in MEDDEV 2.7/1 revision 4. Therefore, any hypothesis that the reliability of the results may be compromised due to protocol adjustment or violation must be rejected and the results must be considered to be valid.

According to this protocol, the results of the comparison require that the null hypothesis, that the cuffs differ in respect to technical, clinical or biological core oscillometric blood pressure measurement technology in the two cuffs, is rejected. Therefore, it must be concluded, that the oscillometric blood pressure measurement technology used in the Style 1 Cuff is equivalent to that used in the Style 2 Cuff. Furthermore, the results of any

validations carried out this oscillometric blood pressure measurement technology, irrespective of which of the two cuffs is used during the validation procedure, must be applied equally to the this technology of both cuffs.

The comparative result proved that the devices are identical with respect to oscillometric measurement. This means that they are the same except in the colour of an accessory, the front deco cover.

It should be noted that SELVAS does not distinguish between the cuffs and the terms “Style 1” and “Style 2” are used only for the purposes of the equivalence check. As the deco ring is part of the cuff, it was necessary to prove equivalence and this was a neutral nomenclature.

Conclusion

The protocol is designed to test the null hypothesis that the devices are different with respect to a particular blood pressure measurement technology and the hypothesis

Recommendations

The SELVAS ACCUNIQ BP503 has been proven to be equivalent to the SELVAS ACCUNIQ BP500 for blood pressure measurement, meaning that the results of all validations of this technology on either device must be applied to the other device. The SELVAS ACCUNIQ BP500 has been proven to be accurate according to the ISO 81060-2:2018 and ISO 81060-2:2018/Amd 1:2020 protocol requirements⁶. Therefore, as the results of this validation must also be applied to the SELVAS ACCUNIQ BP503, and it must be concluded, that the oscillometric blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 monitor is

must be rejected if the respective no difference is found.

As the protocol was followed strictly, any hypothesis that the reliability of the results may be compromised due to protocol adjustment or violation must also be rejected.

Therefore, as all core measurement items used in the measurement of blood pressure by oscillometry were either identical or, in the case of the cuffs, equivalent, there is no option but to reject the null hypothesis and conclude that the oscillometric blood pressure measurement technology used in the SELVAS ACCUNIQ BP503 is equivalent to that used in the SELVAS ACCUNIQ BP500 and vice versa. Therefore, it is imperative that the results of any assessment of this technology carried out using one of these devices must also be applied to the same technology in the other device.

Both cuff styles were found to be identical, as regards blood pressure measurement, with the only difference being the colour of the front deco cover.

accurate for blood pressure measurement in adults, when the device is used according to manufacturer instructions.

Certification Decision

The SELVAS ACCUNIQ BP503, is certified by Medaval Ltd., for blood pressure measurement, in adults, as, proven by equivalence, the technology fulfilled the conditions required for a pass in a validation study carried out in accordance with the requirements of the ISO 81060-2:2018 and ISO 81060-2:2018/Amd 1:2020 standard.

Date of Approval: 06 February 2024.

References

1. European Commission – Health Technology and Cosmetics. MEDDEV 2.7/1 revision 4: Guidelines on Medical Devices – Clinical Evaluation: A Guide for Manufacturers and Notified Bodies under

Directives 93/42/EEC and 90/385/EEC. Brussels, Belgium: European Commission; June 2016 (65 p). Available from: <http://ec.europa.eu/DocsRoom/documents/17522/attachments/1/translations/>

2. The European Parliament and the Council of the European Union. Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC (Text with EEA relevance). Official Journal of the European Union. 2017 May 05;60(L 117):1-175. Available from: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX%3A32017R0745>.
3. ISO/TC 121/SC 3. ISO 81060-2:2018, Non-invasive Sphygmomanometers – Part 2: Clinical Investigation of Automated Measurement Type. Geneva, Switzerland: ISO; 2018 Nov 23.
4. Association for the Advancement of Medical Instrumentation, American National Standards Institute, International Organization for Standardization. AAMI/ANSI/ISO 81060-2:2019, Non-invasive Sphygmomanometers – Part 2: Clinical Investigation of Automated Measurement Type. Arlington, VA, USA: AAMI; 2019 Aug 20.
5. ISO/TC 121/SC 3. ISO 81060-2:2018/Amd 1:2020, Non-invasive sphygmomanometers — Part 2: Clinical investigation of intermittent automated measurement type — AMENDMENT 1. Geneva, Switzerland: ISO; 2020 Jan 08.
6. Medaval Ltd. Full report on the assessment of the blood pressure measurement technology used in the SELVAS ACCUNIQ BP500 upper arm blood pressure monitor, as validated according to the ISO 81060-2:2018 and ISO 81060-2:2018/Amd 1:2020 standard. Medical Device Assessment. 2024 Feb 06;2024(2401FR) 27 p.



SELVAS ACCUNIQ BP503 v BP500 Comparative-Equivalence

Summary of Differences

Item Category	Item Description	Reference Device	Test Device	Comparison
Identity	Primary Device Name	ACCUNIQ BP500	ACCUNIQ BP503	Equivalent on both devices
Identity	All device identities	BP500	BP503	Equivalent on both devices
Accessory	Button Description(s)	Start/Stop: Indigo Acrylonitrile Butadiene Styrene (ABS) Emergency Stop: Red ABS Available in English or Korean	Start/Stop: Green Acrylonitrile Butadiene Styrene (ABS) Emergency Stop: Red ABS Available in English or Korean	Equivalent on both devices
Core (Clinical)	Cuff List	Style 1	Style 2	Equivalent on both devices

Comparison of Primary Details

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Primary Details					
Device Identification					
Identity	Primary Device Name	ACCUNIQ BP500	ACCUNIQ BP503	EL	I
Identity	All device identities	BP500	BP503	EL	I
Identity	All Measurement Modes	Oscillometric	Oscillometric	EP	I
Identity	Defined Measurement Mode	Oscillometric	Oscillometric	EP	I
Manufacturers					
Identity	Branding Company	ACCUNIQ	ACCUNIQ	EP	I
Identity	Distributor			EP	I
Identity	Own Brand Labeller	Not applicable	Not applicable	XB	I
Identity	Original Equipment Manufacturer	SELVAS Healthcare, Inc. 155 Sinseong-ro, Yuseong-gu, Daejeon, Republic of Korea	SELVAS Healthcare, Inc. 155 Sinseong-ro, Yuseong-gu, Daejeon, Republic of Korea	XB	I
Identity	Regulation Manufacturer	Not applicable	Not applicable	XB	I
Identity	Sole Manufacturer	SELVAS Healthcare, Inc. 155 Sinseong-ro, Yuseong-gu, Daejeon, Republic of Korea	SELVAS Healthcare, Inc. 155 Sinseong-ro, Yuseong-gu, Daejeon, Republic of Korea	EP	I
Identity	Other role	Not applicable	Not applicable	XB	I
Documentation	Contact	elliott.k.kim@SELVASHc.com, 82-42-879-3026	elliott.k.kim@SELVASHc.com, 82-42-879-3026	--	--
Primary Descriptors					
Feature	Measuring Functions	Blood Pressure	Blood Pressure	EP	F
Feature	Primary Client Use	Use as a public facility	Use as a public facility	EP	F
Documentation	Validation Publications	None	None	--	--
Core (Clinical)	Measurement Site	Upper Arm	Upper Arm	EP	C
Feature	Measurement Occurrence	Single measurement	Single measurement	EP	F
Documentation	Availability	Available Currently	Available Currently	--	--

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Accessibility	Optional voiced instructions and results	Optional voiced instructions and results	EP	F
Feature	Voiced Languages	Korean / English	Korean / English	EP	F
Documentation	Warranty	0	0	--	A
Files Supplied					
Documentation	User Manual Supplied	Yes	Yes	--	-
Documentation	Service Manual Supplied	Yes	Yes	--	-
Documentation	Specifications Supplied	Yes	Yes	--	-
Documentation	Device Image Supplied			--	-
Documentation	Display Image Supplied			--	-
Documentation	Standards-Compliance supplied	X	None	--	-
Documentation	Pre-clinical studies supplied	None	None	--	-

Comparison of Standard Device Details

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Standard Device Hardware					
<i>Casing</i>					
Accessory	Length	502 mm	502 mm	EP	A
Accessory	Width	450 mm	450 mm	EP	A
Accessory	Height	279 mm	279 mm	EP	A
Accessory	Weight (with batteries)	4800 g	4800 g	EP	A
Accessory	Number of Screens	1	1	EP	A
Accessory	Screen Type	Segment LCD	Segment LCD	EP	A
Accessory	Screen Width	130 mm	130 mm	EP	A
Accessory	Screen Height	90 mm	90 mm	EP	A
Accessory	Screen Backlight	Yes	Yes	EP	A
Accessory	Adjustable Font Size	Not applicable	Not applicable	XB	A
<i>Climate</i>					
Core (Technical)	Minimum Storage Temperature	-10 °C	-10 °C	EP	C
Core (Technical)	Maximum Storage Temperature	+60 °C	+60 °C	EP	C
Core (Technical)	Minimum Operating Temperature	+10 °C	+10 °C	EP	C
Core (Technical)	Maximum Operating Temperature	+40 °C	+40 °C	EP	C
Core (Technical)	Minimum Storage Humidity	0 %RH	0 %RH	EP	C
Core (Technical)	Maximum Storage Humidity	94 %RH	94 %RH	EP	C
Core (Technical)	Non-condensing Storage Humidity	Yes	Yes	EP	C
Core (Technical)	Minimum Operating Humidity	15 %RH	15 %RH	EP	C
Core (Technical)	Maximum Operating Humidity	85 %RH	85 %RH	EP	C
Core (Technical)	Non-condensing Operating Humidity	Yes	Yes	EP	C
Core (Technical)	Minimum Storage Atmospheric Pressure	700 hPa	700 hPa	EP	C
Core (Technical)	Maximum Storage Atmospheric Pressure	1060 hPa	1060 hPa	EP	C
Core (Technical)	Minimum Operating Atmospheric Pressure	700 hPa	700 hPa	EP	C
Core (Technical)	Maximum Operating Atmospheric Pressure	1060 hPa	1060 hPa	EP	C
Core (Technical)	Approximate Maximum Altitude	≈3000 m	≈3000 m	EP	C
<i>Power</i>					
Accessory	Battery Type	No batteries used	No batteries used	EB	A
Accessory	Battery Size	Not applicable	Not applicable	XB	A
Accessory	Battery Details	Not applicable	Not applicable	XB	A
Accessory	Battery Quantity	0	0	EP	A

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Accessory	Battery Life (# measurements)	Not applicable	Not applicable	EP	A
Accessory	Rechargeable battery use	Rechargeable batteries not permitted	Rechargeable batteries not permitted	EB	A
Accessory	AC Adapter Provision	Required – Device only operates from mains	Required – Device only operates from mains	EP	A
Core (Technical)	AC Adapter Number(s)	BRIDGEPOWER BPM060S12F14	BRIDGEPOWER BPM060S12F14	EP	C
Accessory	Automatic Power On	No automatic power on	No automatic power on	EB	A
Accessory	Automatic Power Off	No automatic power off	No automatic power off	EB	A
<i>Communication</i>					
Accessory	Communication Port	USB and RS232	USB and RS232	EP	A
Accessory	Cable Provided	Yes	Yes	EP	A
<i>Audio</i>					
Feature	Voice Memo Recorder	Not provided	Not provided	EB	F
<i>Accessories</i>					
Accessory	Storage/Carrying Case	Not applicable – not intended to be portable	Not applicable – not intended to be portable	XB	A
Accessory	Lid	Not applicable – not intended to be closed	Not applicable – not intended to be closed	XB	A
Accessory	Desk mount facilities	Not applicable	Not applicable	XB	A
Accessory	Wall mount facilities	Not applicable	Not applicable	XB	A
Accessory	Mobile mount facilities	Not applicable	Not applicable	XB	A
Accessory	Pouch	Not applicable	Not applicable	XB	A
Accessory	Belt	Not applicable	Not applicable	XB	A
Accessory	Belt Clip	Not applicable	Not applicable	XB	A
Accessory	Shoulder Straps	Not applicable	Not applicable	XB	A
Accessory	Printer	Integrated as part of device	Integrated as part of device	EP	A
Accessory	Card Holder	Not applicable	Not applicable	XB	A
<i>Appearance</i>					
Accessory	Case Description	White/Grey Available in English and/or Korean	White/Grey Available in English and/or Korean	EP	A
Accessory	Button Description(s)	Start/Stop: Indigo Acrylonitrile Butadiene Styrene (ABS) Emergency Stop: Red ABS Available in English or Korean	Start/Stop: Green Acrylonitrile Butadiene Styrene (ABS) Emergency Stop: Red ABS Available in English or Korean	EL	A
Accessory	Other Description(s)	Not applicable	Not applicable	XB	A
Standard Device Firmware					
<i>Algorithm</i>					
Core (Technical)	Firmware Name and Version	BP500.EN.1.0.00 (Only to distinguish model number)	BP503.EN.1.0.00 (Only to distinguish model number)	EP	C
Standard Device Software					
<i>Memory</i>					
Feature	Number of Memory Locations per User/Zone	0	0	EP	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Number of Users/Zones	0	0	EP	F
Feature	Non-memory use (Guest mode)	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Method of Clearing Memory	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Accessory	Memory card slot	Not provided	Not provided	EB	A
Feature	Date Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Time Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Error Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Facility to mark unique results	Not Applicable (No memory facility)	Not Applicable (No memory facility)	XB	F
<i>Procedure</i>					
Feature	Value shown before measurement	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Average displayed on measurement completion	Not Applicable (No memory facility)	Not Applicable (No memory facility)	XB	F
<i>Measurement</i>					
Core (Technical)	Error Codes	M01 M02 M09 M15 M16 M18 D01 D06 D12 D23 D24	M01 M02 M09 M15 M16 M18 D01 D06 D12 D23 D24	EP	C
<i>Analysis</i>					
Feature	Software Use	No proprietary software provided	No proprietary software provided	EB	F
Feature	Proprietary Software Name	Not applicable	Not applicable	XB	F
<i>Non-Medical Extra Features</i>					
Accessory	Clock	Not provided	Not provided	EB	A
Accessory	Alarm	Not provided	Not provided	EB	A
Accessory	Radio	Not provided	Not provided	EB	A
Accessory	Ambient Temperature	Not provided	Not provided	EB	A
Standard Device Features					
<i>Summary</i>					
Documentation	Features Summary	This device is an electronic device used to measure blood pressure in a non-invasive way outside the body. The cuff is automatically pressurized and systolic and diastolic blood pressure and heart rate are measured and displayed as results.	This device is an electronic device used to measure blood pressure in a non-invasive way outside the body. The cuff is automatically pressurized and systolic and diastolic blood pressure and heart rate are measured and displayed as results.	--	--
Documentation	Items not listed	None	None	--	--
Standard Screen and Audio Indicators					
<i>Measurement</i>					
<i>After Measurement</i>					
Feature	Measurement Unit(s)	Shown	Shown	EP	F
Feature	Average – Overall	Function not provided	Function not provided	EB	F
Feature	Plot	No plot provided	No plot provided	EB	F
<i>Error Indicators</i>					
Feature	Measurement Error	“Error” text with code or number	“Error” text with code or number	EP	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
<i>Timestamp</i>					
Feature	Date and Time	Date and Time shown	Date and Time shown	EP	F
Feature	Time Format	24-hour and 12-hour clocks	24-hour and 12-hour clocks	EP	F
<i>Markers</i>					
Feature	Event – Medication	Function not provided	Function not provided	EB	F
<i>Memory</i>					
Feature	Value from Memory	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Recorded Measurement Timestamp	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Memory Location Number	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Memory Locations Used	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Memory Full	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Memory Zone Name	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Delete memory	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
<i>Non-Measurement</i>					
<i>Power</i>					
Feature	Battery Symbol	Not applicable (No battery)	Not applicable (No battery)	XB	F
Feature	Battery Charging Indicator	Not applicable (No charging)	Not applicable (No charging)	XB	F
Feature	Voltage Check	Function not provided	Function not provided	EB	F
Feature	AC Adapter Symbol	Function not provided	Function not provided	EB	F
Feature	Power Error Symbol	Error code	Error code	EP	F
Feature	Start	Function not provided	Function not provided	EB	F
Feature	Stop	Function not provided	Function not provided	EB	F
<i>Communication</i>					
Feature	Reminder to Transfer Data	Function not provided	Function not provided	EB	F
Feature	Device Connected	Special Icon	Special Icon	EP	F
Feature	Transmitting Data	Function not provided	Function not provided	EB	F
Feature	Transmission Successful	Function not provided	Function not provided	EB	F
Feature	Transmission Unsuccessful	Function not provided	Function not provided	EB	F
Feature	Signal out-of-range	Function not provided	Function not provided	EB	F
Feature	PC Link	Reuse of 7-segment characters	Reuse of 7-segment characters	EP	F
<i>Settings</i>					
Feature	Settings	Function not provided	Function not provided	EB	F
Feature	Initialisation	Function not provided	Function not provided	EB	F
Feature	Hide measurement display option	Not provided	Not provided	EB	F
<i>Audio Indicators</i>					
Feature	Sound Volume	Seven levels	Seven levels	EP	F
Feature	Measurement Value	Optional Voiced Indicator	Optional Voiced Indicator	EP	F
Feature	Memory Value	None	None	EB	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Statistics Value	None	None	EB	F
Feature	Measurement Reminder	None	None	EB	F
Feature	Measurement Complete	Optional Voiced Indicator	Optional Voiced Indicator	EP	F
Feature	Measurement Error	Optional Voiced Indicator	Optional Voiced Indicator	EP	F
Feature	Voice Recorder	Function not provided	Function not provided	EB	F
Display					
Feature	Screen Background Colour(s)	Black	Black	EP	A
Feature	Screen Font Colour(s)	White	White	EP	A
Feature	Screen Language(s)	None	None	EB	A
Standard Buttons and Switches					
<i>Power</i>					
Feature	Power On	Power	Power	EP	F
Feature	Power Off	Power	Power	EP	F
Feature	Start	Start/Stop	Start/Stop	EP	F
Feature	Stop	Start/Stop	Start/Stop	EP	F
<i>Up/Down</i>					
Feature	Increase value	Up	Up	EP	F
Feature	Decrease value	Down	Down	EP	F
Feature	Previous value	Not applicable	Not applicable	XB	F
Feature	Next value	Not applicable	Not applicable	XB	F
Feature	Increase volume	Set and Up	Set and Up	EP	F
Feature	Decrease volume	Down and Set	Down and Set	EP	F
<i>Memory</i>					
Feature	Memory mode	Not applicable	Not applicable	XB	F
Feature	Memory bank selection	Not applicable	Not applicable	XB	F
Feature	Delete memory	Not applicable	Not applicable	XB	F
Feature	Delete last measurement	Not applicable	Not applicable	XB	F
<i>Date and Time</i>					
Feature	Date and Time settings	Down, Set and Up	Down, Set and Up	EP	F
Feature	Alarm settings	Not applicable	Not applicable	XB	F
Feature	Date and Time announcement	Not applicable	Not applicable	XB	F
<i>Statistics</i>					
Feature	Show Average	Not applicable	Not applicable	XB	F
Feature	Show Alternative Average	Not applicable	Not applicable	XB	F
Feature	Show Plot	Not applicable	Not applicable	XB	F
<i>Settings</i>					
Feature	Settings	Set	Set	EP	F
Feature	Confirm	Not applicable	Not applicable	XB	F
Feature	Measurement unit settings	Not applicable	Not applicable	XB	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Change language	User	User	EP	F

Comparison of BPM-Specific Details

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
BPM Device Hardware					
<i>Cuffs</i>					
Core (Clinical)	Cuff List	Style 1	Style 2	EL	C
<i>Sensors</i>					
Core (Technical)	Number of pressure sensors	1	1	EP	C
Core (Technical)	Pressure sensor type	Strain gauge	Strain gauge	EP	C
Core (Technical)	Pressure sensor model number(s)/code(s)	ADP1131	ADP1131	EP	C
Core (Technical)	Pressure sensor cross-check	Not provided	Not provided	EB	C
Documentation	Pressure sensor details supplied	Yes	Yes	--	--
Core (Technical)	Positioning sensor type	Capacitive position sensor	Capacitive position sensor	EP	C
Core (Technical)	Positioning sensor model number(s)/code(s)	BS812A-1	BS812A-1	EP	C
Documentation	Positioning sensor details supplied	Yes	Yes	--	--
Core (Technical)	Cuff-Wrapping Sensor	Not provided	Not provided	EB	C
Core (Technical)	ECG sensor	No sensor	No sensor	EB	C
Documentation	ECG sensor details supplied	Not applicable	Not applicable	--	--
Core (Technical)	Korotkoff-sound sensor	No sensor	No sensor	EB	C
Documentation	Korotkoff-sound sensor details supplied	Not applicable	Not applicable	--	--
Core (Technical)	Activity sensor	Not provided	Not provided	EB	C
Documentation	Activity sensor details supplied	Not applicable	Not applicable	--	--
<i>Signal Processing</i>					
Core (Technical)	Amplifier model number(s)/code(s)	LM324DR	LM324DR	EP	C
Core (Technical)	Analogue filter model number(s)/code(s)	Not applicable	Not applicable	XB	C
Core (Technical)	Analogue-to-digital convertor model number(s)/code(s)	ADC within Micom (ATSAM4S16CA)	ADC within Micom (ATSAM4S16CA)	EP	C
Core (Technical)	Pressure Sampling rate	1000 Hz	1000 Hz	EP	C
<i>Pneumatic Hardware</i>					
Core (Technical)	Pneumatic pump model number/code	P54A-0001R (abr P54A01R) or RFP45J-0002R (abr RFP45J02R)	P54A-0001R (abr P54A01R) or RFP45J-0002R (abr RFP45J02R)	EP	C
Documentation	Pneumatic pump details supplied	Yes	Yes	--	--

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Core (Technical)	Exhaust valve model number/code	KSV15C-6I	KSV15C-6I	EP	C
Documentation	Exhaust valve details supplied	Yes	Yes	--	-
Core (Technical)	Safety Release Valve	Not provided	Not provided	EB	C
<i>Accessories</i>					
Accessory	Cuff Holder	Cuff integrated into device	Cuff integrated into device	XB	A
BPM Device Software					
<i>Memory</i>					
Feature	SBP Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	DBP Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	PR Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	MAP Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Generic Event Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Medication Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Atrial Fibrillation Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Arrhythmia/IHB Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	BP Grade Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Body Movement Code Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Bed and Rising Times Stored	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
<i>Ranges</i>					
Core (Technical)	Maximum Pressure (Upper Limit of Upper Technical Alarm Condition Range)	300 mmHg	300 mmHg	EP	C
Core (Technical)	Upper Limit of Rated Range	280 mmHg	280 mmHg	EP	C
Core (Technical)	Lower Limit of Rated Range	30 mmHg	30 mmHg	EP	C
Core (Technical)	Minimum Pressure (Lower Limit of Low Technical Alarm Condition Range)	0 mmHg	0 mmHg	EP	C
Core (Technical)	Maximum SBP	280 mmHg	280 mmHg	EP	C
Core (Technical)	Minimum SBP	60 mmHg	60 mmHg	EP	C
Core (Technical)	Maximum DBP	200 mmHg	200 mmHg	EP	C
Core (Technical)	Minimum DBP	30 mmHg	30 mmHg	EP	C
Core (Technical)	Maximum PP	250 mmHg	250 mmHg	EP	C
Core (Technical)	Minimum PP	15 mmHg	15 mmHg	EP	C
Core (Technical)	Maximum PR	240 bpm	240 bpm	EP	C
Core (Technical)	Minimum PR	30 bpm	30 bpm	EP	C
<i>Specified Accuracy</i>					
Core (Technical)	Specified BP Accuracy (±)	2 mmHg	2 mmHg	EP	C
Core (Technical)	Specified PR Accuracy (±)	1.5 %	1.5 %	EP	C

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
BPM Device Firmware					
<i>Analysis</i>					
Feature	Overall arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last overall 3-meas arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last overall 7-day arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last overall ESH arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Overall median	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Overall morning arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last morning 3-meas arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last morning 7-day arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Overall evening arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last evening 3-meas arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	Last evening 7-day arithmetic mean	Not applicable (No memory facility)	Not applicable (No memory facility)	XB	F
Feature	BP Classification	Not provided	Not provided	EB	F
Feature	Pulse Classification	Not provided	Not provided	EB	F
Feature	Measurement-Target BP difference	Not provided	Not provided	EB	F
BPM Device Features					
<i>Technical</i>					
Core (Technical)	Operation Method	Oscillometry: automatic during deflation	Oscillometry: automatic during deflation	EP	C
Feature	Single Measurements	Yes	Yes	EP	F
Feature	Double Measurement	Not provided	Not provided	EB	F
Feature	Triple Measurement	Not provided	Not provided	EB	F
Feature	Measurements at ESH-recommended times	Not provided	Not provided	EB	F
Feature	ABPM Measurement Occurrences	Device does not support ABPM	Device does not support ABPM	XB	F
Feature	ABPM duration	Not applicable (Not ABPM)	Not applicable (Not ABPM)	XB	F
Core (Technical)	Continuous Measurements	Not provided	Not provided	EB	C
Feature	Continuous mode frequency	Not applicable (No continuous mode)	Not applicable (No continuous mode)	XB	F
Core (Technical)	ECG triggered measurements (ECG used to confirm Korotkoff sounds)	Not provided	Not provided	EB	C
Feature	Measurement Interval Set	Not applicable	Not applicable	XB	F
Feature	Measurement Times Set	Not applicable	Not applicable	XB	F
Core (Clinical)	Repeat measurement	No automatic repeat provided	No automatic repeat provided	EB	C
Feature	Suspend Multiple/ABPM measurements	Not applicable (No multiple measurements)	Not applicable (No multiple measurements)	XB	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
<i>Calibration</i>					
Feature	Calibration Mode	Not provided	Not provided	EB	F
Core (Technical)	Recommended Calibration Intervals	Not applicable	Not applicable	EP	C
Feature	Recalibrate Reminder	Not provided	Not provided	EB	F
<i>Special Measurements</i>					
Core (Technical)	Test Measurements	None	None	EB	C
Core (Clinical)	Measurements triggered by ECG event	Not provided	Not provided	EB	C
Feature	Measurements triggered by posture	Not provided	Not provided	EB	F
Feature	Measurements triggered by activity	Not provided	Not provided	EB	F
Feature	Manually triggered measurements (ABPM only)	Not provided	Not provided	EB	F
<i>Procedure</i>					
Core (Technical)	Positioning Check	Yes	Yes	EP	C
Core (Technical)	Zero Pressure Check	Not provided	Not provided	EB	C
Core (Technical)	Inflation Method	Automatic after manual-initiated start	Automatic after manual-initiated start	EP	C
Core (Technical)	Inflation Rate	Depends on upper arm thickness and expected SBP value	Depends on upper arm thickness and expected SBP value	EP	C
Core (Technical)	Inflation Target	Measurement during inflation – SBP dependent	Measurement during inflation – SBP dependent	EP	C
Core (Technical)	Inflation Target Value(s)	Not applicable	Not applicable	XB	C
Core (Technical)	Deflation Method	Logical control (drop after SBP)	Logical control (drop after SBP)	EP	C
Core (Technical)	Deflation Rate	4	4	EP	C
<i>Measurement</i>					
Feature	Measurement Units	mmHg only	mmHg only	EP	F
Core (Clinical)	Systolic Blood Pressure (SBP)	On Screen and Report	On Screen and Report	EP	C
Core (Clinical)	Diastolic Blood Pressure (DBP)	On Screen and Report	On Screen and Report	EP	C
Core (Clinical)	Pulse Rate (PR)	On Screen and Report	On Screen and Report	EP	C
Core (Clinical)	Measured Mean Arterial Pressure (MAP)	Not provided	Not provided	EB	C
Feature	Estimated Mean Arterial Pressure (MAP)	Optionally on Report only	Optionally on Report only	EP	F
Feature	Pulse Pressure (PP)	Optionally on Report only	Optionally on Report only	EP	F
Feature	Pressure Rate Product (PRP)	Optionally on Report only	Optionally on Report only	EP	F
Feature	Left and right arm pressures	Not provided	Not provided	EB	F
Core (Clinical)	Central Aortic Pressures	Not provided	Not provided	EB	C
Feature	Ankle-Brachial Index	Not provided	Not provided	EB	F
Core (Technical)	Technical Alarm Condition	Included in standard out-of-range error	Included in standard out-of-range error	EP	C

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Arterial Stiffness Index	Not provided	Not provided	EB	F
Feature	Posture	Not provided	Not provided	EB	F
Feature	Activity Level	Not provided	Not provided	EB	F
<i>ISO Certification - Manual Devices</i>					
Documentation	ISO-81060-1 Certification (Manual devices only)	Not applicable (automatic)	Not applicable (automatic)	--	--
Documentation	Date of ISO Certification (Manual devices only)	Not applicable	Not applicable	--	--
BPM Screen and Audio Indicators					
Measurement					
<i>Before Measurement</i>					
Feature	Measurement Mode	Function not provided	Function not provided	EB	F
Feature	Arm positioning indicator	Special Icon	Special Icon	EP	F
Feature	Wrist positioning indicator	Function not provided	Function not provided	EB	F
Feature	Posture indicator	Function not provided	Function not provided	EB	F
Feature	Cuff wrapping indicator	Function not provided	Function not provided	EB	F
Feature	Left/Right Limb Selection	Function not provided	Function not provided	EB	F
Feature	Inflation Target Selection	Not applicable for intended device use	Not applicable for intended device use	XB	F
Feature	Threshold Selection	Function not provided	Function not provided	EB	F
Feature	Memory zone	Function not provided	Function not provided	EB	F
<i>During Measurement</i>					
Feature	Inflation	Reuse of 7-segment characters	Reuse of 7-segment characters	EP	F
Feature	Deflation	Reuse of 7-segment characters	Reuse of 7-segment characters	EP	F
Feature	Heartbeat Indicator	Special Icon	Special Icon	EP	F
Feature	Pressure	Digital value	Digital value	EP	F
<i>After Measurement</i>					
Feature	SBP	Always shown	Always shown	EP	F
Feature	DBP	Always shown	Always shown	EP	F
Feature	Measured MAP	Not provided	Not provided	EB	F
Feature	PR	Always shown	Always shown	EP	F
Feature	Posture	Function not provided	Function not provided	EB	F
Feature	Activity Level	Function not provided	Function not provided	EB	F
Feature	Pulse Wave Pattern	Available on printout	Available on printout	EP	F
<i>Derived Values</i>					
Feature	PP	Available on printout	Available on printout	EP	F
Feature	Estimated MAP ((SBP + 2 × DBP) / 3)	Available on printout	Available on printout	EP	F
Feature	Pressure Rate Product (PRP = SBP × PR)	Available on printout	Available on printout	EP	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Average – Morning	Function not provided	Function not provided	EB	F
Feature	Average – Evening	Function not provided	Function not provided	EB	F
Feature	BP Classification	Classification not indicated	Classification not indicated	EB	F
Feature	Pulse Classification	Function not provided	Function not provided	EB	F
Feature	Arrhythmias	Function not provided	Function not provided	EB	F
Feature	Haemodynamic Stability	Function not provided	Function not provided	EB	F
Feature	Inter-Arm Difference	Function not provided	Function not provided	EB	F
Feature	Arterial Stiffness	Function not provided	Function not provided	EB	F
Feature	Visit doctor	Function not provided	Function not provided	EB	F
<i>Error Indicators</i>					
Feature	Body-Movement Error	Function not provided	Function not provided	EB	F
Feature	Air leak/Cuff Connection Error	Function not provided	Function not provided	EB	F
Feature	Ambient temperature Error	Function not provided	Function not provided	EB	F
Feature	Measurement Reliability Error	Function not provided	Function not provided	EB	F
Feature	Measurement being repeated	Function not provided	Function not provided	EB	F
<i>Markers</i>					
Feature	Event – Generic	Function not provided	Function not provided	EB	F
Feature	Event – Bed Time & Rising Time	Function not provided	Function not provided	EB	F
<i>Memory</i>					
Feature	Pulse Rate from Memory	Not provided	Not provided	EB	F
<i>Special Measurements</i>					
Feature	Test Measurements	Function not provided	Function not provided	EB	F
Feature	Test Successful	Function not provided	Function not provided	EB	F
Non-Measurement					
<i>Settings</i>					
Feature	Button Lock	Function not provided	Function not provided	EB	F
Audio Indicators					
Feature	Measurement imminent	None	None	EB	F
Feature	Pulse signal detected	Not applicable (No sound)	Not applicable (No sound)	XB	F
BPM Buttons and Switches					
<i>Safety</i>					
Feature	Immediate Exhaust	Emergency	Emergency	EP	F
<i>Mode</i>					
Feature	Mode: Single Measurement	Start/Stop	Start/Stop	EP	F
Feature	Mode: Double Measurement	Not applicable	Not applicable	XB	F
Feature	Mode: Triple Measurement	Not applicable	Not applicable	XB	F
Feature	Mode: Home Measurement	Not applicable	Not applicable	XB	F
Feature	Mode: ABPM	Not applicable	Not applicable	XB	F

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Feature	Mode: Nocturnal repeated measurements	Not applicable	Not applicable	XB	F
Feature	Mode: Diagnostic	Not applicable	Not applicable	XB	F
Feature	Mode: ESH controlled	Not applicable	Not applicable	XB	F
Feature	Mode: Auscultation/Manual	Not applicable	Not applicable	XB	F
Feature	Mode: Automatic	Not applicable	Not applicable	XB	F
<i>Settings</i>					
Feature	Manual Threshold Selection	Not applicable	Not applicable	XB	F
Feature	Left/Right Limb Selection	Not applicable	Not applicable	XB	F
Feature	Inflation Target Selection	Not applicable	Not applicable	XB	F
<i>Event</i>					
Feature	Event: Generic	Not applicable	Not applicable	XB	F
Feature	Event: Medication	Not applicable	Not applicable	XB	F
Feature	Event: Bed Time	Not applicable	Not applicable	XB	F
Feature	Event: Rising Time	Not applicable	Not applicable	XB	F

Summation of Comparisons

Category		Identity	Core	Feature	Accessory	Core Breakdown			Total	
Comparison	Code	I	C	F	A	Biological	Clinical	Technical		
Identical provision on both devices	EP	5	55	48	18	0	4	51	126	
Equivalent provision on both devices	EL	2	1	0	1	0	1	0	4	
Not provided on either device	EB	0	14	68	10	0	4	10	92	
Similar-level provisions on both devices	BL	0	0	0	0	0	0	0	0	
Different provisions on each device	BP	0	0	0	0	0	0	0	0	
Provided on Reference but not on Test Device	RP	0	0	0	0	0	0	0	0	
Better provision on Reference Device	RL	0	0	0	0	0	0	0	0	
Provided on Test but not on Reference Device	TP	0	0	0	0	0	0	0	0	
Better provision on Test Device	TL	0	0	0	0	0	0	0	0	
Not comparable	XP	0	0	0	0	0	0	0	0	
Not applicable for this device functionality	XB	4	2	77	15	0	0	2	98	
Value missing for Reference Device	MR	0	0	0	0	0	0	0	0	
Value missing for Test Device	MT	0	0	0	0	0	0	0	0	
Values missing for both devices	MB	0	0	0	0	0	0	0	0	
Check if provision is identical or similar	QL	0	0	0	0	0	0	0	0	
Check free text entries	QP	0	0	0	0	0	0	0	0	
Totals	Formulae									
Equal items	$\Sigma EP + \Sigma EL + \Sigma EB$	En	7	70	116	29	0	9	61	222
Different	$Rn + Tn + Bn$	Dn	0	0	0	0	0	0	0	0
Favouring Reference	$\Sigma RP + \Sigma RL$	Rn	0	0	0	0	0	0	0	0
Favouring Test	$\Sigma TP + \Sigma TL$	Tn	0	0	0	0	0	0	0	0
Favouring Neither	$\Sigma BP + \Sigma BL$	Bn	0	0	0	0	0	0	0	0
Missing	$\Sigma MR + \Sigma MT + \Sigma MB$	Mn	0	0	0	0	0	0	0	0
Queries	$\Sigma QP + \Sigma QL$	Qn	0	0	0	0	0	0	0	0
Not Compared	$\Sigma XP + \Sigma XB$	Xn	4	2	77	15	0	0	2	98
Overall Sum	$En + Dn + Mn + Qn + Xn$	Sn	11	72	193	44	0	9	63	320
Applicable Sum	$Sn - Xn$	An	7	70	116	29	0	9	61	222

Equivalence and Comparative Analysis

Equivalence Analysis																	
Description	Formulae	Core															
Not Equivalent	$Dn-C > 0$	False															
Incomplete Core	$Mn-C > 0$	False															
Core Queries	$Qn-C > 0$	False															
Equivalence Proven	Reject if any Criteria Test results are TRUE	True															
Comparative Analysis Completion Check																	
Description	Formulae		<table border="1"> <thead> <tr> <th>Feature</th> <th>Accessory</th> </tr> </thead> <tbody> <tr> <td>193 (100.0%) Complete</td> <td>44 (100.0%) Complete</td> </tr> <tr> <td>193 (100.0%) Complete</td> <td>44 (100.0%) Complete</td> </tr> <tr> <td>193 (100.0%) Complete</td> <td>44 (100.0%) Complete</td> </tr> <tr> <td colspan="2" style="text-align: center;">(100.0%) Complete</td> </tr> <tr> <td colspan="2" style="text-align: center;">False (0 outstanding)</td> </tr> <tr> <td colspan="2" style="text-align: center;">True</td> </tr> </tbody> </table>	Feature	Accessory	193 (100.0%) Complete	44 (100.0%) Complete	193 (100.0%) Complete	44 (100.0%) Complete	193 (100.0%) Complete	44 (100.0%) Complete	(100.0%) Complete		False (0 outstanding)		True	
Feature	Accessory																
193 (100.0%) Complete	44 (100.0%) Complete																
193 (100.0%) Complete	44 (100.0%) Complete																
193 (100.0%) Complete	44 (100.0%) Complete																
(100.0%) Complete																	
False (0 outstanding)																	
True																	
Sufficiency for Reference Device	$(Sn - Qn - MR - MB) / Sn \geq 66\%, 88\%$																
Sufficiency for Test Device	$(Sn - Qn - MT - MB) / Sn \geq 66\%, 88\%$																
Sufficiency for Comparison	$(Sn - Qn - Mn) / Sn \geq 66\%, 88\%$																
Overall Sufficiency	Minimum of above $\geq 66\%, 88\%$																
Details Queries	$Qn-F + Qn-A > 0$																
Details finalised	Reject if Queries or Insufficient																
Comparative Analysis Results Check																	
Option	Formulae		Feature + Accessory														
Identical	$Dn-F = 0 \wedge Dn-A = 0$		True														
Common	$Dn-F = 0 \wedge Dn-A > 0 \vee Rn-F = 0 \wedge Tn-F = 0 \wedge En-F \geq 9 \times Bn-F$		True														
Simpler	$Rn-F > 0 \wedge Tn-F = 0 \vee Tn-F > 0 \wedge Rn-F \geq 3 \times Tn-F$		False														
Superior	$Tn-F > 0 \wedge Rn-F = 0 \vee Rn-F > 0 \wedge Tn-F \geq 3 \times Rn-F$		False														
Diverse	$0 < Rn-F < 3 \times Tn-F \wedge 0 < Tn-F < 3 \times Rn-F \wedge Rn-F + Tn-F \geq 3 \times B-Fn$		False														
Common-Diverse	$Rn-F = 0 \wedge Tn-F = 0 \wedge En-F < 9 \times Bn-F \vee 0 < Rn-F < 3 \times Tn-F \wedge 0 < Tn-F < 3 \times Rn-F \wedge Rn-F + Tn-F < 3 \times Bn-F$		False														
Comparative Result			<p style="text-align: center;">Identical</p> <p style="text-align: center;"><i>This denotes the situation where the test device has identical functions/accessories to the reference device.</i></p>														



Cuff Comparative-Equivalence

Summary of Differences

Item Category	Item Description	Reference Device	Test Device	Comparison
Identity	Primary Cuff Name	SELVAS Style 1	SELVAS Style 1	Equivalent on both devices
Accessory	Relevant Accessories	Indigo deco cover	Grey deco cover	Equivalent on both devices

Comparison of Details

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
Cuff Details					
<i>Cuff Identification</i>					
Identity	Primary Cuff Name	SELVAS Style 1	SELVAS Style 2	EL	I
Identity	All cuff identities	Integrated	Integrated	EP	I
Identity	Generic Name	Universal	Universal	EP	I
Identity	Manufacturer	As Device	As Device	EP	I
<i>Descriptors</i>					
<i>Size</i>					
Core (Technical)	Minimum Circumference	200 mm	200	EP	C
Core (Technical)	Maximum Circumference	400 mm	400	EP	C
Core (Technical)	Bladder Length	146 mm	146	EP	C
Core (Technical)	Bladder Width	158 mm	158	EP	C
<i>Cuff Details</i>					
Core (Technical)	Cuff Style	Rigid cylindrical upper-arm cuff	Rigid cylindrical upper-arm cuff	EP	C
Core (Technical)	Bladder Style	Integrated cuff and bladder	Integrated cuff and bladder	EP	C
Core (Technical)	Cuff Fastenings	Integrated cuff without fastenings	Integrated cuff without fastenings	EP	C
Core (Biological)	Outer Material	Spandex	Spandex	EP	C
Feature	Connector	Not applicable - Integrated cuff	Not applicable - Integrated cuff	EP	F
<i>Accessories</i>					
Accessory	Relevant Accessories	Indigo deco cover	Grey deco cover	EL	A
<i>Microphone Details</i>					
Core (Technical)	Microphones	No microphone	No microphone	XB	C
Core (Technical)	Microphone model number(s)/code(s)	Not Applicable	Not Applicable	EB	C
<i>Provision</i>					
Documentation	Default Provision	Integrated as part of device	Integrated as part of device	--	--
Documentation	Availability	Available Currently	Available Currently	--	--

Item Category	Item Description	Reference Device	Test Device	Comparison	Category
<i>Validation and Certification</i>					
Documentation	ISO-81060-1 Certification (Cuffs for manual devices only)	Not applicable (automatic)	Not applicable (automatic)	--	-
Documentation	Date of ISO Certification (Cuffs for manual devices only)	Not applicable	Not applicable	--	-
<i>Files Supplied</i>					
Documentation	Cuff Image Supplied			--	-
Documentation	Microphone details	Not Applicable	Not Applicable	--	-

Summation of Comparisons

Category		Identity	Core	Feature	Accessory	Core Breakdown			Total
Comparison	Code	I	C	F	A	Biological	Clinical	Technical	
Identical provision on both devices	EP	3	8	1	0	1	0	7	12
Equivalent provision on both devices	EL	1	0	0	1	0	0	0	2
Not provided on either device	EB	0	1	0	0	0	0	1	1
Similar-level provisions on both devices	BL	0	0	0	0	0	0	0	0
Different provisions on each device	BP	0	0	0	0	0	0	0	0
Provided on Reference but not on Test Device	RP	0	0	0	0	0	0	0	0
Better provision on Reference Device	RL	0	0	0	0	0	0	0	0
Provided on Test but not on Reference Device	TP	0	0	0	0	0	0	0	0
Better provision on Test Device	TL	0	0	0	0	0	0	0	0
Not comparable	XP	0	0	0	0	0	0	0	0
Not applicable for this device functionality	XB	0	1	0	0	0	0	1	1
Value missing for Reference Device	MR	0	0	0	0	0	0	0	0
Value missing for Test Device	MT	0	0	0	0	0	0	0	0
Values missing for both devices	MB	0	0	0	0	0	0	0	0
Check if provision is identical or similar	QL	0	0	0	0	0	0	0	0
Check free text entries	QP	0	0	0	0	0	0	0	0
Totals	Formulae								
Equal items	$\Sigma EP + \Sigma EL + \Sigma EB$	En	4	9	1	1	0	8	15
Different	$Rn + Tn + Bn$	Dn	0	0	0	0	0	0	0
Favouring Reference	$\Sigma RP + \Sigma RL$	Rn	0	0	0	0	0	0	0
Favouring Test	$\Sigma TP + \Sigma TL$	Tn	0	0	0	0	0	0	0
Favouring Neither	$\Sigma BP + \Sigma BL$	Bn	0	0	0	0	0	0	0
Missing	$\Sigma MR + \Sigma MT + \Sigma MB$	Mn	0	0	0	0	0	0	0
Queries	$\Sigma QP + \Sigma QL$	Qn	0	0	0	0	0	0	0
Not Compared	$\Sigma XP + \Sigma XB$	Xn	0	1	0	0	0	1	1
Overall Sum	$En + Dn + Mn + Qn + Xn$	Sn	4	10	1	1	0	9	16
Applicable Sum	$Sn - Xn$	An	4	9	1	1	0	8	15

Equivalence and Comparative Analysis

Equivalence Analysis																	
Description	Formulae	Core															
Not Equivalent	$Dn-C > 0$	False															
Incomplete Core	$Mn-C > 0$	False															
Core Queries	$Qn-C > 0$	False															
Equivalence Proven	Reject if any Criteria Test results are TRUE	True															
Comparative Analysis Completion Check																	
Description	Formulae		<table border="1"> <thead> <tr> <th>Feature</th> <th>Accessory</th> </tr> </thead> <tbody> <tr> <td>1 (100.0%) Complete</td> <td>1 (100.0%) Complete</td> </tr> <tr> <td>1 (100.0%) Complete</td> <td>1 (100.0%) Complete</td> </tr> <tr> <td>1 (100.0%) Complete</td> <td>1 (100.0%) Complete</td> </tr> <tr> <td colspan="2">(100.0%) Complete</td> </tr> <tr> <td colspan="2">False (0 outstanding)</td> </tr> <tr> <td colspan="2">True</td> </tr> </tbody> </table>	Feature	Accessory	1 (100.0%) Complete	1 (100.0%) Complete	1 (100.0%) Complete	1 (100.0%) Complete	1 (100.0%) Complete	1 (100.0%) Complete	(100.0%) Complete		False (0 outstanding)		True	
Feature	Accessory																
1 (100.0%) Complete	1 (100.0%) Complete																
1 (100.0%) Complete	1 (100.0%) Complete																
1 (100.0%) Complete	1 (100.0%) Complete																
(100.0%) Complete																	
False (0 outstanding)																	
True																	
Sufficiency for Reference Device	$(Sn - Qn - MR - MB) / Sn \geq 66\%, 88\%$																
Sufficiency for Test Device	$(Sn - Qn - MT - MB) / Sn \geq 66\%, 88\%$																
Sufficiency for Comparison	$(Sn - Qn - Mn) / Sn \geq 66\%, 88\%$																
Overall Sufficiency	Minimum of above $\geq 66\%, 88\%$																
Details Queries	$Qn-F + Qn-A > 0$																
Details finalised	Reject if Queries or Insufficient																
Comparative Analysis Results Check																	
Option	Formulae		Feature + Accessory														
Identical	$Dn-F = 0 \wedge Dn-A = 0$		True														
Common	$Dn-F = 0 \wedge Dn-A > 0 \vee Rn-F = 0 \wedge Tn-F = 0 \wedge En-F \geq 9 \times Bn-F$		True														
Simpler	$Rn-F > 0 \wedge Tn-F = 0 \vee Tn-F > 0 \wedge Rn-F \geq 3 \times Tn-F$		False														
Superior	$Tn-F > 0 \wedge Rn-F = 0 \vee Rn-F > 0 \wedge Tn-F \geq 3 \times Rn-F$		False														
Diverse	$0 < Rn-F < 3 \times Tn-F \wedge 0 < Tn-F < 3 \times Rn-F \wedge Rn-F + Tn-F \geq 3 \times B-Fn$		False														
Common-Diverse	$Rn-F = 0 \wedge Tn-F = 0 \wedge En-F < 9 \times Bn-F \vee 0 < Rn-F < 3 \times Tn-F \wedge 0 < Tn-F < 3 \times Rn-F \wedge Rn-F + Tn-F < 3 \times Bn-F$		False														
Comparative Result			Identical <i>This denotes the situation where the test device has identical functions/accessories to the reference device</i>														

Appendix 1 – Medaval Comparative-Equivalence Procedure

Cover Page

Medaval Comparative-Equivalence Specifications

Version 4



Medaval Comparative-Equivalence Procedure

Authors

Roland Asmar	Yutaka Imai	Andrew Shennan
Neil Atkins	Martin Myers	Jan Staessen
Alejandro de la Sierra	Eoin O'Brien	George S. Stergiou
Peter de Leeuw	Gbenga Ogedegbe	Martin Turner
Eamon Dolan	Takayoshi Ohkubo	Paolo Verdecchia
Przemyslaw Guzik	Paolo Palatini	Bernard Waeber
Geoffrey A. Head	Gianfranco Parati	Jiguang Wang

Effective 23 March 2017

Summary

Medaval Comparative-Equivalence Specifications

Version 4

Summary

This document describes equivalence for medical devices in accordance with MEDDEV 2.7/1 revision 4.

- In compliance with Council Directive 93/42/EEC as amended by directive 2007/47/EC
- Used to compare devices described under Medaval Device Registration with Core measurement-critical items indicated as Technical, Clinical or Biological, in accordance with MEDDEV 2.7/1 revision 4
- Identifies how each item is compared to corresponding item on the paired device
- Standard critical core items compared to ensure equality or equivalence, in accordance with MEDDEV 2.7/1 revision 4.
- Identity, Features and Accessories indicated, in Device Registration, compared to identify device differences in accordance with MEDDEV 2.7/1 revision 4.
- Provides summary comparison for equivalent devices
- Identifies "Families" of devices where each pair are equivalent to each other
- Results provided in Clinical Evaluation Reports prepared in accordance with MEDDEV 2.7/1 revision 4.
- Evaluations conducted by renowned experts in Blood Pressure Measurement with proven and published experience in protocol development and validation.

Equivalence and Validation

- Equivalence is independent of validation.
- Validation is only of one functionality, normally the the measurement technology. The particular device used in a validation can be any that uses that technology.
- Validation of a measurement technology is therefore applied to all devices within a family i.e. all those with proven to be equivalent for that technology.
- Existing validations are applied immediately. Subsequent validations are applied once published.
- Equivalence of cuffs means that where one cuff was proved accurate with the technology, all equivalent cuffs can be used also.

Appendix 2 –Comparative-Equivalence of all Devices in Family

The SELVAS ACCUNIQ BP503 and the SELVAS ACCUNIQ BP500 are just two devices in a family of devices that have been proven to be mutually equivalent. The other devices are the SELVAS ACCUNIQ BP501, the SELVAS ACCUNIQ BP600, the SELVAS ACCUNIQ BP650, the SELVAS ACCUNIQ BP651 and the SELVAS ACCUNIQ BP653.













In summary, the model number digits describe the differences between the devices. The first digit is either 5 (button material ABS) or 6 (button material acrylic). The second digit is either 0 (no

BP classification provided) or 5 (BP classification and a separate arm positioning indicator) and the third digit is either 0 (thermal printer and derived values only provided on a printout, indigo start stop button and indigo deco cover), 1 (no printer and no derived values, green start stop button and grey deco cover) or 3 (thermal printer and the printed derived values, green start stop button and grey cuff deco cover).

Proof of the other equivalences are contained in separate reports and summarised below.

Comparative Result	Test Device						
	BP500	BP501	BP503	BP600	BP650	BP651	BP653
BP500	Same	Simpler	Identical	Identical	Superior	Diverse	Superior
BP501	Superior	Same	Superior	Superior	Superior	Superior	Superior
BP503	Identical	Simpler	Same	Identical	Superior	Diverse	Superior
BP600	Identical	Simpler	Identical	Same	Superior	Diverse	Superior
BP650	Simpler	Simpler	Simpler	Simpler	Same	Simpler	Identical
BP651	Diverse	Simpler	Diverse	Diverse	Superior	Same	Superior
BP653	Simpler	Simpler	Simpler	Simpler	Identical	Simpler	Same

Legend: Identical means equivalent and no effective differences in features or accessories; Diverse means equivalent and each device has features not available on the other; Superior means equivalent and more features; Simpler means equivalent and fewer features

Item Description	ACCUNIQ BP500	ACCUNIQ BP501	ACCUNIQ BP503	ACCUNIQ BP600	ACCUNIQ BP650	ACCUNIQ BP651	ACCUNIQ BP653
Device Image							
Display							
Integrated Printer	Yes	No	Yes	Yes	Yes	No	Yes
Availability of printed PP, Estimated MAP, PRP and plot	Yes	No	Yes	Yes	Yes	No	Yes
Button Material	ABS	ABS	ABS	Acrylic	Acrylic	Acrylic	Acrylic
Start/Stop Button Colour	Indigo	Green	Green	Indigo	Indigo	Green	Green
Deco Cover Colour	Indigo	Grey	Grey	Indigo	Indigo	Grey	Grey
BP Classification	No	No	No	No	Yes	Yes	Yes
Arm positioning indicator	Heart	Heart	Heart	Heart	Special Icon	Special Icon	Special Icon